

ANNEXURE NO. SABS 1

TO CERTIFICATE OF APPROVAL NO. 842

IN ACCORDANCE WITH THE PROVISIONS OF SECTION 18(2)(b) OF THE TRADE METROLOGY ACT (ACT 77 OF 1973), THE DIRECTOR OF TRADE METROLOGY AFTER CONSULTATION WITH THE METROLOGY ADVISORY COMMITTEE, HEREBY CERTIFIES THAT THE PATTERN OF THE MODIFIED INSTRUMENTS DESCRIBED HEREUNDER, WHICH CONFORM TO ALL DOCUMENTS DEPOSITED WITH THE SOUTH AFRICAN BUREAU OF STANDARDS FOR APPROVAL PURPOSES, MEETS THE REQUIREMENTS OF THE TRADE METROLOGY ACT AND MAY BE USED FOR PRESCRIBED PURPOSES AFTER DUE CONSIDERATION OF ANY LIMITATIONS OR CONDITIONS IMPOSED BY THE PATTERN DESCRIPTION:

The undermentioned "Kent" mechanical, positive displacement, oscillating piston, domestic water meters. These meters may incorporate an internal non return valve and may be fitted with a magnetic reed switch for supplementary external interrogation.

Nominal Bore	Model	Lengths	q_p	Class	Position of Installation
15mm	KSM	114 mm	1,5 m ³ /h	B or C	Horizontal or Vertical
15mm	PSM	114 mm & 165 mm	1,5 m ³ /h	B or C	Horizontal or Vertical
20 mm	PSM	165 mm	2,5 m ³ /h	B or C	Horizontal or Vertical
25 mm	PSM	198 mm	3,5 m ³ /h	B or C	Horizontal or Vertical

The above meters were submitted for re-approval according to the requirements of specification SABS 1529-1:1994.

NOTES:


- 1. This annexure replaces the annexure issued on 26 March 1997 which did not include the 15 mm KSM and 25 mm PSM Models.*
- 2. This annexure relates to the suitability of the equipment for use for prescribed purposes only in respect of its metrological characteristics. It does not constitute or imply any guarantee as to the safety of the equipment in use for trade or otherwise.*

SUBMITTED BY:

ABB Kent Measurement (Pty) Ltd
Industria, JOHANNESBURG

26 August 1998

DATE



DIRECTOR: TRADE METROLOGY
South African Bureau of Standards
Private Bag X191
0001 PRETORIA
REPUBLIC OF SOUTH AFRICA